

Safe Work Procedure (SWP)

Name of Task: SCH 2A OnOff Toilet w FI Lift	
Description of Task: Transfer client on/off toilet or commode using two person assist and mechanical floor lift.	
Position/Job: HCA, ISW, RA, nurse	Department/Unit: WRHA Home Care
Equipment and/or Tools Required:	Personal Protective Equipment (PPE) Required:
Mechanical floor lift (appropriate weight capacity)	Appropriate footwear and clothing for staff
Sling appropriate for client	
Commode	
May be used: electric bed, bedrail	

<u>Potential Hazards:</u> Forceful movements, awkward/sustained positions, infection transmission, violent and/or aggressive behavior

Signs and symptoms of a musculoskeletal injury (MSI) can include pain, burning, swelling, stiffness, numbness/tingling, and/or loss of movement or strength in a body part. Report these to your supervisor.

Considerations:

- Client must be assessed for ability and potential concerns in using this lift. Do not proceed with transfer and contact supervisor if there are changes in client's ability.
- If behavior interferes with care, additional procedures and/or care planning may be required. Report to supervisor.
- Client must be able to:
 - Maintain upright sitting balance independently on toilet/commode
- If applying sling requires more than minimal assistance by each staff (e.g., rolling client), additional procedures and/or equipment may be required (e.g., using sliders to insert sling)

Note: Prior to each use, ensure that the lift is in proper working order and that the battery is sufficiently charged. Inspect the slings for signs of damage prior to use.

Note: Mechanical lifts are designed for short distance transfers in the same room (e.g., bed to commode).

Note: Client should not be left unattended while attached to lift and sitting on the toilet/commode.

Steps to be taken to complete task safely:

Set-up:

1

- Establish staff roles: The staff most familiar with the client and the procedure takes the lead. Lead staff is responsible for directing the task (communicating with client and second staff); second staff is responsible for watching for safety and assisting as directed by lead. Note: second staff may be a caregiver or staff with medical restrictions
- Where applicable, position commode with enough room to complete transfer and maneuver lift.
 Remove footrests and apply brakes if wheeled commode.
- If starting on bed, adjust bed to waist height of shortest staff wherever possible.
- Adjust client's clothing prior to transfer.
- Ensure that there is enough space to move the lift and that the flooring is appropriate (e.g., hardwood, vinyl, tile, low pile carpet).
- Both staff stand on opposite sides of bed/chair/commode wherever possible

2 Sling application:

If client is sitting:

- Ask client to lean forward. Insert sling behind client until the bottom edge reaches seat pan of the
 chair. If client requires minimal assist to lean forward, staff stand at side of client in a wide base of
 support, hips and knees bent and back straight. Staff places one hand on client's upper back and
 the other hand in front of client's shoulder and shift weight to side to lean client forward.
- Ask client to lift leg slightly to insert leg strap. Gently pull leg strap around inner thigh and ensure bottom edge of sling reaches client's tailbone. Be careful not to shear client's skin.
- One staff may hold onto client's shin below knee to slightly raise client's thigh from seated surface while second staff inserts leg strap. If available, place a slider under client's thigh to position leg strap.

If client is lying:

- Ask client to roll to one side and place center of sling along spine of client. If it is difficult to roll client, use sliders.
- Ensure bottom of sling is positioned at client's tailbone; top of sling is positioned at/above client's shoulders (as per sling type). If client requires assist to roll, staff stands on either side of client with wide base of support and one foot in front of the other. Lead staff pushes/rolls client towards second

WRHA Home Care Version 1.0

	person by shifting weight up and forward. Second staff assists to guide and maintain client in place. With one side of the sling draped over client, lead staff tucks half of sling under client.
	Ask or assist client to roll to opposite side to pull sling through.
3	Attach sling to lift:
	 Open floor lift base as wide as possible. Insert under bed frame or around chair/commode/toilet base. Brakes do not need to be applied for the duration of the transfer.
	• Lead staff lowers the hanger bar over the client's chest (in bed) or in front of client (in sitting) so the loops of the sling can be easily attached to the lift.
	 Leg straps are positioned as appropriate for sling type (e.g., looped through and crossed). Follow any special instructions provided.
	 Staffs securely attach appropriate loops based on client's size and desired position on toilet/commode (i.e., sitting). Most commonly attach longer loop at leg straps and shorter loop at shoulder straps. Ensure the same level/loop colors are used on the left and right side of the body at the each location (i.e., shoulders and legs).
4	Lift client:
	Ask client to cross arms to prepare for transfer.
	 Lead staff uses the controller to raise client just enough to clear bed or seated surface/armrests (may require lowering bed surface).
	Ensure that all loops remain securely attached to the hanger bar and that client is properly
	positioned and comfortable in sling before continuing with the lift/transfer.
5	Move client to toilet/commode:
	 With elbows tucked in and wrist straight, lead staff grasps onto handles of lift. With one foot in front of the other, lead staff pulls lift away and guides the floor lift and client towards receiving surface.
	Use weight shifting technique to move lift. If client is on bed, second staff uses two hands to move
	client's legs off the bed and turns client to face the mast of the lift. If lift is difficult to turn, staff may
	stand beside the lift and hold onto one handle while using a foot to push on the base.
	 Second staff moves the commode/wheelchair under the client and ensures brakes are applied.
6	Lower client onto commode/toilet:
	Ensure client is centered over toilet/commode
	Lead staff lowers client using the controller, ensuring hanger bar does not contact client's head.
	Option #1: Lead staff stands onside of lift in front of client and second staff stands behind client/commode. Both staff stand with one foot in front of the other and hips and knees bent and back straight. Lead staff gently pushes on leg straps and/or client's shins while second staff holds onto handles of sling to guide the client back onto the commode. Both staff use weight shift technique (one forward, one back).
	Option #2: Both staff stands on either side of lift in front of client and gently push on leg straps
	and/or client's shins to guide client back. Both staff use weight shift technique forward.
	Note: When lowering onto commode, it may tip slightly back. This corrects when client comes close to
	seat.
7	Sling removal:
	Once client is positioned correctly, unhook sling from lift.
	Remove one leg strap at a time by folding it under itself and pull towards outside of chair. Ask client
	to assist by lifting thigh slightly if able. One staff may hold onto client's shin below knee to slightly
	raise client's thigh from seated surface while second staff removes leg strap.
	Reverse procedure to transfer client back to original surface.
	Responsibilities
Manage	rs/Supervisors or Designate: Ensure all duties are performed in accordance to training, established health

<u>Managers/Supervisors or Designate</u>: Ensure all duties are performed in accordance to training, established health and safety regulations/guidelines, policies and procedures (e.g. following safe work procedures)

<u>Staff performing task</u>: Perform task in accordance to training, established health and safety regulations/guidelines, policies and procedures (e.g. following safe work procedures). Notify Manager/Supervisor or designates of all occurrences/incidents, injuries, illnesses or safety and health concerns which are likely to harm themselves, coworkers, or any others who enter the premises.

Completion and Review

This safe work procedure will be reviewed any time the task, equipment or materials change and at a minimum of every three years from the last revision date

Approved by: Pending approval by PWHSWG and PHRLC

Created by: Home Care Safe Client Handling Committee

Date: February 2019

WRHA Home Care Version 1.0